



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSL Z540-1-1994

CAL-RITE CORPORATION
1665 Quincy Avenue, Suite 103
Naperville, IL 60540
Nathan Hathaway Phone: 630 355 1522

CALIBRATION

Valid To: November 30, 2026

Certificate Number: 0866.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations^{1,7}:

I. Dimensional

Parameter/Equipment	Range	CMC ^{2,4,5} ($\pm$)	Comments
Calipers ³	Up to 6 in (> 6 to 20) in (> 20 to 60) in	(69 + 0.6R) μ in (200 + 0.6R) μ in (460 + 0.6R) μ in	Gage blocks
Displacement Length Indicators ³	Up to 4 in (> 4 to 12) in (> 12 to 60) in	7.2L + 31 μ in 12L + 8.8 μ in (460 + 0.6R) μ in	Gage blocks, digital micrometer head
Extensometers/COD Gage/Deflectometers ³			ASTM E83 & ISO 9513;
Travel 2 in & below	Up to 0.02 in (> 0.02 to 2.0) in	19 μ in 37 μ in	Gage blocks, extensometer calibrator; uncertainties listed in displacement
Travel > 2 to 25 in	(> 2 to 25) in	69L + 33 μ in	Gage blocks
Gage Length	Up to 8 in	0.0012 in	Caliper
Micrometers ³ –	Up to 6 in (> 6 to 20) in	(71 + 0.6R) μ in (200 + 0.6R) μ in	Gage blocks

Parameter/Equipment	Range	CMC ^{2, 4, 5} ($\pm$)	Comments
Microscopes ³ – Vision Systems	Up to 1 in Up to 25.4 mm	52 μ m 1.7 μ m	ASTM E1951 stage micrometers

II. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2, 4, 5, 6} ($\pm$)	Comments
DC Voltage – Generate Electrical Calibration of Load Indicators	$\pm$ (0 to 4.4) mV/V	0.000 12 mV/V	Precision simulator
Electrical Simulation of Thermocouple Indicating Devices	Type N (-200 to 1300) °C Type J (-210 to 1200) °C Type K (-270 to 1372) °C Type T (-250 to 400) °C	0.60 °C 0.60 °C 0.60 °C 0.60 °C	Additel 222A
Alignment System Calibration	100 μ e 500 μ e 1000 μ e 2000 μ e 3000 μ e	1.4 μ e 1.5 μ e 1.6 μ e 1.9 μ e 2.2 μ e	Strain indicator calibrator

III. Mechanical

Parameter/Equipment	Range	CMC ^{2, 4, 5} ($\pm$)	Comments
Scales & Balances ³	(1 to 200) mg 300 mg 500 mg 1 g to 20 kg (0.001 to 0.1) lb (> 0.1 to 2) lb (> 2 to 600) lb	0.012 mg 0.012 mg 0.012 mg 0.029 % 0.029 % 0.029 % 0.029 %	Class 1 Certified weights
Mass Measure	(1 to 200) mg 300 mg 500 mg 1 g 2 g 3 g 5 g 10 g 20 g 30 g 50 g 100 g 200 g 205 g 300 g 500 g 1 kg 2 kg 3 kg 5 kg 5.1 kg 10 kg 11.8 kg 20 kg 25 kg 30 kg 32.1 kg	0.016 mg 0.026 mg 0.026 mg 0.027 mg 0.027 mg 0.029 mg 0.030 mg 0.037 mg 0.043 mg 0.068 mg 0.071 mg 0.12 mg 0.21 mg 0.22 mg 0.60 mg 0.58 mg 1.8 mg 6.7 mg 6.0 mg 9.5 mg 9.5 mg 47 mg 47 mg 50 mg 55 mg 121 mg 122 mg	Mass comparator, mass standards

Parameter/Equipment	Range	CMC ^{2, 4, 5} (±)	Comments
Force ³ – Measure			
Tension & Compression by Deadweight with Local Gravity & Air Bouyancy Corrections	(0.01 to 20) lbf	0.04 %	ASTM E4 & ISO 7500-1
Comparison to Load Cell in Compression	(1 to 1 000 000) lbf	0.25 %	ASTM E4 within the Class A working range, ISO 7500-1 within the Class 0.5 or 1 working range
Comparison to Load Cell in Tension	(1 to 250 000) lbf	0.25 %	
Force3 – Measuring Equipment (Tension & Compression)			Force calibrations include:
Calibration by Deadweight with Local Gravity & Air Buoyancy Corrections	(0.1 to 1000) lbf	0.05 %	ASTM E74: within Class AA working range
Comparison to Load Cell in Compression / Tension	(10 to 1000) lbf	0.05 %	
	(1000 to 10 000) lbf	0.01 %	
	(10 000 to 100 000) lbf	0.01 %	ISO 376: within Class 00, 0.5 & 1 working ranges
Comparison to Load Cell in Compression / Tension	(10 to 1000) lbf	0.05 %	
	(1000 to 10 000) lbf	0.05 %	
	(10 000 to 100 000) lbf	0.05 %	

Parameter/Equipment	Range	CMC ^{2, 4, 5} (±)	Comments
Verification of Test Frames ³ –			
Specimen Alignment (50 to 10 000) lbf	(100 to 3000) µe	1.0 % + 5 µe	ASTM E1012 alignment calibration system
Crosshead Displacement	(0 to 2) in (2 to 25) in	250 µin 0.23 %	ASTM E2309 digital indicator cable extension transducer
Crosshead Speed	(0.005 to 20) in/min	0.33 %	ASTM E2658 cable extension transducer with stopwatch
Strain Rate	(0.0005 to 0.015) in/in/min	0.00012 in/in/min	Calibrated extensometer within ASTM Class B1 range & stopwatch
Load Rate	(50 to 600 000) lb/min	0.32 %	Class A Load cell & stopwatch
Platen Flatness	0.001 in	56 µin	Feeler Gauge & Straight Edge
Pressure Gauges ³ / Transducer	(0 to 2000) psi (2000 to 40 000) psi	0.70 psi 2.4 psi	Pressure calibrator

Parameter/Equipment	Range	CMC ^{2, 4, 5} (±)	Comments
Hardness Testers ³ – Rockwell	HRA: (20 to 65) HRA (70 to 78) HRA (80 to 84) HRA HRBW: (40 to 59) HRBW (60 to 79) HRBW (80 to 100) HRBW HRC: (20 to 30) HRC (35 to 55) HRC (60 to 65) HRC HREW: (70 to 79) HREW (84 to 90) HREW (93 to 100) HREW HRFW: (60 to 75) HRFW (80 to 90) HRFW (94 to 100) HRFW HRGW: (30 to 50) HRGW (55 to 75) HRGW (80 to 94) HRGW HRHW: (80 to 94) HRHW (94 to 100)HRHW HRKW: (40 to 60) HRKW (65 to 80) HRKW (85 to 100) HRKW HRMW: Low High	0.44 HRA 0.34 HRA 0.27 HRA 0.69 HRBW 0.58 HRBW 0.60 HRBW 0.41 HRC 0.40 HRC 0.36 HRC 0.55 HREW 0.61 HREW 0.53 HREW 0.52 HRFW 0.51 HRFW 0.59 HRFW 0.60 HRGW 0.49 HRGW 0.49 HRGW 0.59 HRHW 0.44 HRHW 0.54 HRKW 0.62 HRKW 0.47 HRKW 0.55 HRMW 0.44 HRMW	Indirect verification per ASTM E18

Parameter/Equipment	Range	CMC ^{2, 4, 5} ($\pm$)	Comments
Hardness Testers ³ – Rockwell (cont)	HRLW: Low High HRRW: (100 to 120) HRRW HR15TW: (74 to 80) HR15TW (81 to 86) HR15TW (87 to 93) HR15TW HR30TW: (43 to 56) HR30TW (57 to 69) HR30TW (70 to 83) HR30TW HR45TW: (13 to 32) HR45TW (33 to 52) HR45TW (53 to 73) HR45TW HR15N: (70 to 77) HR15N (78 to 88) HR15N (90 to 92) HR15N HR30N: (42 to 50) HR30N (55 to 73) HR30N (77 to 82) HR30N HR45N: (20 to 31) HR45N (37 to 61) HR45N (66 to 72) HR45N HR15WW: Low High HR30WW: Low High	0.55 HRLW 0.45 HRLW 0.58 HRRW 0.62 HR15TW 0.40 HR15TW 0.40 HR15TW 0.50 HR30TW 0.46 HR30TW 0.43 HR30TW 0.49 HR45TW 0.53 HR45TW 0.57 HR45TW 0.48 HR15N 0.40 HR15N 0.40 HR15N 0.54 HR30N 0.57 HR30N 0.51 HR30N 0.45 HR45N 0.48 HR45N 0.50 HR45N 0.54 HR15WW 0.50 HR15WW 0.78 HR30WW 0.62 HR30WW	Indirect verification per ASTM E18

Parameter/Equipment	Range	CMC ^{2, 4, 5} ($\pm$)	Comments
Hardness Testers ³ – Rockwell (cont)	HR15YW: Low High HR15TS: (74 to 80) HR15TS (81 to 86) HR15TS (87 to 93) HR15TS HR30TS: (43 to 56) HR30TS (57 to 69) HR30TS (70 to 83) HR30TS HRBS: (40 to 59) HR30BS (60 to 79) HR30BS (80 to 100) HR30BS	0.52 HR15YW 0.52 HR15YW 0.61 HR15TS 0.61 HR15TS 0.41 HR15TS 0.50 HR30TS 0.44 HR30TS 0.43 HR30TS 0.71 HRBS 0.69 HRBS 0.58 HRBS	Indirect verification per ASTM E18
Hardness Testers – Brinell			
HBW 10/3000	(95 to 650) HBW	6.5 HBW	Indirect verification method per ISO 6506-2, ASTM E10
HBW 10/3000	(96 to 650) HBW	7.3 HBW	
HBW 10/1500	(48 to 327) HBW	3.8 HBW	
HBW 10/1000	(32 to 218) HBW	2.0 HBW	
HBW 5/750	(96 to 650) HBW	9.6 HBW	
HBW 10/500	(16 to 109) HBW	1.9 HBW	
HBW 10/250	(32 to 218) HBW	1.8 HBW	
HBW 2.5/187.5	(96 to 650) HBW	9.9 HBW	
HBW 2.5/62.5	(32 to 218) HBW	6.3 HBW	

Parameter/Equipment	Range	CMC ^{2, 4, 5} (±)	Comments
Hardness Testers – Brinell (cont)			Indirect verification method per ISO 6506-2, ASTM E10
Direct Verification of Brinell Hardness Testers			
Direct Verification of the Test Force	(62.5, 187.5, 250, 500, 750, 1000, 1500, 3000) kgf	0.25 %	Load cells within ASTM E4 Class A working range
Verification of the Brinell Scope ⁵	(0 to 10) mm Type A Type B	0.007 mm 0.06 mm	Stage micrometer
Verification of test cycle	Up to 15 sec	0.08 sec	Stopwatch
Leeb Hardness Testers ³	(517 to 759) HLD	9.6 HLD	ASTM A956
Microindentation Hardness Testers (Knoop & Vickers) ³ –			
For Loads Less than 1000 g	(100 to 250) HK (250 to 650) HK > 650 HK (100 to 240) HV (240 to 600) HV > 600 HV	1.0 % HK 1.0 % HK 1.0 % HK 0.8 % HV 0.9 % HV 0.9 % HV	Indirect verification method per ASTM E92, ISO 6507-2
For Loads Greater than 1000 g	Low HV Mid HV High HV	1.3 % HV 1.3 % HV 1.3 % HV	
X-Y Stage	(0 to 1) inch (0 to 25.4) mm	65 µin 1.4 µm	Stage micrometer

Parameter/Equipment	Range	CMC ^{2, 4, 5} ($\pm$)	Comments
Indirect Verification of Charpy Impact Testers ³ Specimen Evaluation	Low Energy High Energy Super High Energy Low Energy Mid Energy High Energy	1.1 ft/lb 2.5 ft/lb 4.1 ft/lb 1.2 J 3.5 J 4.9 J	ASTM E23 ISO 148-2
Direct Verification of Charpy Impact Testers ³ Anvil to Anvil Measure	40 mm	0.069 mm	ASTM E23 & ISO 148-2; Caliper & telescope gage
Center Strike Measure	20 mm	0.61 mm	Caliper
Charpy Base Level	0°	0.15°	Precision level
Charpy Bolt Torque	(30 to 150) lbf*ft	4.8 %	Torque wrench

IV. Thermodynamics

Parameter/Equipment	Range	CMC ^{2, 4, 5} ($\pm$)	Comments
Temperature – Measure ³	(-80 to 590) °C (0 to 1000) °C	0.09 °C 1.0 °C 2.0	Fluke 1529 with 5628 PRT Fluke 743B with Type K TC
Humidity – Measure ³	(11 to 95) % RH	0.70 % RH	Vaisala HMP77B

V. Time & Frequency

Parameter/Equipment	Range	CMC ^{2, 4, 5} ($\pm$)	Comments
Timers & Stopwatches ³	(0 to 24) hrs	0.04 sec/24 hrs	Timometer

¹ This laboratory offers commercial calibration service and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ In the statement of CMC, percentages are percent of reading unless otherwise indicated.

⁵ Uncertainties may differ depending on the performance of artifact being measured at the time of calibration and the resolution of the device under test.

⁶ The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.

⁷ This scope meets A2LA's *P112 Flexible Scope Policy*.



Accredited Laboratory

A2LA has accredited

CAL-RITE CORPORATION

Naperville, IL

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCCL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 30th day of September 2024.

A blue ink signature of Mr. Trace McInturff.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 866.01
Valid to November 30, 2026
Revised October 19, 2026

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.